

Schedule

Unitest Instruments Pte Ltd
18 Boon Lay Way
#08-127
Singapore 609966

Certificate No. : LA-2023-0845-C
Issue No. : 1
Date : 03 August 2023
Expiry of Certificate : 02 August 2027
Page : 1 of 20

Scope / Instruments / Location / Method	Range to be Calibrated / Measured Quantities / Frequency	CALIBRATION AND MEASUREMENT CAPABILITY (CMC*)
A ELECTRICAL – DC/LF		
1. Resistance measuring instrument Lab In-House procedure UNI-E001 Rev:0	0 Ω – 12 Ω >12 Ω – 120 Ω >120 Ω – 1.2 kΩ >1.2 kΩ – 12 kΩ >12 kΩ – 120 kΩ >120 kΩ – 1.2 MΩ >1.2 MΩ – 12 MΩ >12 MΩ – 120 MΩ >120 MΩ – 1200 MΩ	0.011 Ω 0.13 % 0.11 % 0.019 % 0.011 % 0.011 % 0.021 % 0.062 % 0.42 %
2. DC Voltage measuring instrument Lab-In-House procedure UNI-E001 Rev:0	0 mV – 120 mV >120 mV – 1.2 V >1.2 V – 12 V >12 V – 120 V >120 V – 1020 V	0.0024 mV 0.0011 % 0.0017 % 0.0024 % 0.0024 %
3. AC Voltage measuring instruments Lab-In-House procedure UNI-E001 Rev:0	1.0 mV to 12 mV 3 Hz to 5 Hz >5 Hz to 10 kHz >10 kHz to 20 kHz >20 kHz to 50 kHz >50 kHz to 100 kHz >100 kHz to 500 kHz	1.6% 1.0 % 0.65 % 0.66 % 1.7 % 3.9 %

Schedule



Certificate No. : LA-2023-0845-C

Issue No. : 01

Date : 03 August 2023

Page : 2 of 20

Scope / Instruments / Location / Method	Range to be Calibrated / Measured Quantities / Frequency	CALIBRATION AND MEASUREMENT CAPABILITY (CMC*)
	<p>12 mV to 120 mV</p> <p>3 Hz – 5 Hz >5 Hz – 10 kHz >10 kHz – 20 kHz >20 kHz – 50 kHz >50 kHz – 100 kHz >100 kHz – 500 kHz</p> <p>120 mV – 1.2 V</p> <p>3 Hz – 5 Hz >5 Hz – 10 Hz >10 Hz – 40 Hz >40 Hz – 20 kHz >20 kHz – 50 kHz >50 kHz – 100 kHz >100 kHz – 500 kHz</p> <p>1.2 V – 12 V</p> <p>3 Hz – 5 Hz >5 Hz – 10 Hz >10 Hz – 40 Hz >40 Hz – 20 kHz >20 kHz – 50 kHz >50 kHz – 100 kHz > 100 kHz – 500 kHz</p> <p>12 V – 120 V</p> <p>>3 Hz – 5 Hz >5 Hz – 10 Hz >10 Hz – 40 Hz >40 Hz – 20 kHz >20 kHz – 50 kHz >50 kHz – 100 kHz</p> <p>120 V – 330 V</p> <p>3 Hz – 5 Hz >5 Hz – 10 Hz >10.1 Hz – 20 kHz >20 kHz – 50 kHz >50 kHz – 100 kHz</p>	<p>0.32 % 0.17 % 0.065 % 0.11 % 0.25 % 1.3 %</p> <p>0.42 % 0.15 % 0.6 % 0.13 % 0.047 % 0.11 % 0.41 %</p> <p>0.33 % 0.5 % 0.6 % 0.018 % 0.035 % 0.081 % 0.25</p> <p>0.32 % 0.50 % 0.045 % 0.092 % 0.036 % 0.082 %</p> <p>0.50 % 0.16 % 0.021 % 0.039 % 0.17 %</p>

Schedule



Certificate No. : LA-2023-0845-C

Issue No. : 01

Date : 03 August 2023

Page : 3 of 20

Scope / Instruments / Location / Method	Range to be Calibrated / Measured Quantities / Frequency	CALIBRATION AND MEASUREMENT CAPABILITY (CMC*)
	330V – 1020 V 3 Hz – 5 Hz >5 Hz – 10 Hz >10 Hz – 10 kHz	0.29 % 0.12 % 0.04 %
4. DC Current measuring instruments		
Lab-In-House procedure UNI-E001 Rev:0	0 μ A – 120 μ A >120 μ A – 1.2 mA >1.2 mA – 12 mA >12 mA – 120 mA >120 mA – 1.2 A >1.2 A – 3.1 A >3.1 A – 12 A >12 A – 30 A	0.021 μ A 0.045 % 0.019 % 0.02 % 0.029 % 0.044 % 0.041 % 0.11 %
5. DC Current Clamp measuring instruments		
Lab-In-House procedure UNI-E002 Rev:0	0 mA – 20 mA 20 mA – 100 mA 0.1 A – 4 A >4 A – 40 A >40 A – 60 A >60 A – 200 A >200 A – 300 A >300 A – 1500A	0.14 mA 0.72 % 1.0 % 0.7 % 0.67 % 0.67 % 0.67 % 0.67 %
6. AC Current measuring instruments		
Lab-In-House procedure UNI-E001 Rev:0	10 μA – 120 μA >3 Hz – 22.5 Hz >22.5 Hz – 45 Hz >45 Hz – 1 kHz >1 kHz – 5 kHz >5 kHz – 10 kHz >10 kHz – 30 kHz 120 μA – 1.2 mA 3 Hz – 22.5 Hz >22.5 Hz – 45 Hz >45 Hz – 1 kHz >1 kHz – 5 kHz >5 kHz – 10 kHz >10 kHz – 30 kHz	0.14 % 0.06 % 0.06 % 0.06 % 0.14 % 3.4 % 0.14 % 0.11 % 0.12 % 0.11 % 0.24 % 4.8 %

Schedule



Certificate No. : LA-2023-0845-C

Issue No. : 01

Date : 03 August 2023

Page : 4 of 20

Scope / Instruments / Location / Method	Range to be Calibrated / Measured Quantities / Frequency	CALIBRATION AND MEASUREMENT CAPABILITY (CMC*)
	<p>1.2 mA – 12 mA 3 Hz – 45 Hz >45 Hz – 1 kHz >1 kHz – 5 kHz >5 kHz – 10 kHz >10 kHz – 30 kHz</p> <p>12 mA – 120 mA 3 Hz – 45 Hz >45 Hz – 1 kHz >1 kHz – 5 kHz >5 kHz – 10 kHz >10 kHz – 30 kHz</p> <p>120 mA – 1.2 A 3 Hz – 45 Hz >45 Hz – 1 kHz >1 kHz – 5 kHz >5 kHz – 10 kHz >10 kHz – 30 kHz</p> <p>1.2 A – 3.1 A 3 Hz – 45 Hz >45 Hz – 1 kHz >1 kHz – 5 kHz >5 kHz – 10 kHz</p> <p>3.1 A – 12 A 3 Hz – 45 Hz >45 Hz – 1 kHz >1 kHz – 5 kHz >5 kHz – 10 kHz</p> <p>12 A – 30.2 A 3 Hz – 45 Hz >45 Hz – 1 kHz >1 kHz – 5 kHz</p>	<p>0.11 % 0.11 % 0.11 % 0.24 % 1.4 %</p> <p>0.11 % 0.06 % 0.018 % 0.24 % 1.4 %</p> <p>0.34 % 0.068 % 0.09 % 0.5 % 0.8 %</p> <p>0.25 % 0.06 % 0.06 % 0.3 %</p> <p>0.12 % 0.05 % 0.07 % 0.29 %</p> <p>0.21 % 0.14 % 0.57 %</p>

Schedule



Certificate No. : LA-2023-0845-C

Issue No. : 01

Date : 03 August 2023

Page : 5 of 20

Scope / Instruments / Location / Method	Range to be Calibrated / Measured Quantities / Frequency	CALIBRATION AND MEASUREMENT CAPABILITY (CMC*)
7. AC Current clamp measuring instrument Lab-In-House procedure UNI-E002 Rev:0	(30 Hz – 60 Hz) 10 mA – 40 mA 40 mA – 400 mA 0.4 A – 4 A >4 A – 40 A >40 A – 60 A >60 A – 200 A >200 A – 300 A >300 A – 1500A	0.76 % 0.70 % 0.67 % 0.68 % 0.72 % 0.69 % 0.68 % 0.67 %
8. Frequency measuring instruments Lab-In-House procedure UNI-E001 Rev:0	0.01 Hz – 119.99 Hz 120.0 Hz – 1199.9 Hz 1.200 kHz – 11.999 kHz 12.00 kHz – 119.99 kHz 120.1 kHz – 1199.9 kHz 1.200 MHz – 2.000 MHz	0.018 % 0.0014 % 0.0011 % 0.00057 % 0.53 % 0.00027 %
9. Capacitance measuring instruments Lab-In-House procedure UNI-E001 Rev:0	0.2 nF – 1.2 nF >1.2 nF – 12 nF >12 nF – 120 nF >120 nF – 1.2 μF >1.2 μF – 12 μF >12 μF – 120 μF >120 μF – 1.2mF >1.2 mF – 12 mF >12 mF – 120 mF	2.3 % 0.6 % 0.4 % 0.4% 0.4 % 0.4 % 0.45 % 0.49 % 0.8 %
10. Inductance measuring instruments Lab-In-House procedure UNI-E001 Rev:0	13 μH – 120 μH >120 μH – 1.2 mH >1.2 mH – 12 mH >12 mH – 120 mH >120 mH – 1.2 H >1.2 H – 12 H >12 H – 120 H	2.0 % 1.0 % 1.0 % 1.0 % 1.0 % 1.0 % 1.1 %

Schedule



Certificate No. : LA-2023-0845-C

Issue No. : 01

Date : 03 August 2023

Page : 6 of 20

Scope / Instruments / Location / Method	Range to be Calibrated / Measured Quantities / Frequency	CALIBRATION AND MEASUREMENT CAPABILITY (CMC*)
11. Resistance sourcing instruments Lab-In-House procedure UNI-E003 Rev:0	0.01 Ω – 1 Ω >1 Ω – 10 Ω >10 Ω – 100 Ω >100 Ω – 1 k Ω >1 k Ω – 10 k Ω >10 k Ω – 100 k Ω >100 k Ω – 1 M Ω >1 M Ω – 10 M Ω >10 M Ω – 100 M Ω >100 M Ω – 1 G Ω >1 G Ω – 10 G Ω	0.16 % 0.0058 % 0.0042 % 0.002% 0.003 % 0.004 % 0.008 % 0.02 % 0.13 % 0.17 % 1.2 %
12. DC Voltage sourcing instruments Lab-In-House procedure UNI-E003 Rev:0	0 mV – 100 mV >0.11 V – 1 V >1 V – 10 V >10 V – 100 V >100 V – 1000 V	0.0021 mV 0.0012 % 0.001 % 0.002 % 0.005 %
13. AC Voltage sourcing instruments Lab-In-House procedure UNI-E003 Rev:0	1 mV – 10 mV 3 Hz – 2 kHz >2 kHz – 10 kHz >10 kHz – 30 kHz >30 kHz – 100 kHz >100 kHz – 300 kHz 10 mV – 100 mV 3 Hz – 2 kHz >2 kHz – 10 kHz >10 kHz – 30 kHz >30 kHz – 100 kHz >100 kHz – 300 kHz	0.26 % 0.26 % 0.27 % 0.53 % 1.9 % 0.57 % 0.57 % 0.07 % 0.19 % 1.0 %

Schedule



Certificate No. : LA-2023-0845-C

Issue No. : 01

Date : 03 August 2023

Page : 7 of 20

Scope / Instruments / Location / Method	Range to be Calibrated / Measured Quantities / Frequency	CALIBRATION AND MEASUREMENT CAPABILITY (CMC*)	
14. DC Current sourcing instruments Lab-In-House procedure UNI-E003 Rev:0	100 mV – 1 V 3Hz – 2 kHz	0.02 %	
	1 V – 10 V 3Hz – 10 kHz >10 kHz – 30 kHz >30 kHz – 100 kHz >100 kHz – 300 kHz	0.03 % 0.05 % 0.14 % 0.66 %	
	10 V – 100 V 3 Hz – 2 kHz >2 kHz – 10 kHz >10 kHz – 30 kHz >30 kHz – 100 kHz >100 kHz – 300 kHz	0.57 % 0.57 % 0.10 % 0.15 % 1.3 %	
	100 V – 1000 V 3 Hz – 2 kHz >2 kHz – 10 kHz >10 kHz – 30 kHz	0.05 % 0.05 % 0.10 %	
	0 μ A – 10 μ A >10 μ A – 100 μ A >100 μ A – 1 mA >1 mA – 10 mA >10 mA – 100 mA >100 mA – 1 A >1 A – 10 A 1>0 A – 30 A	0.0018 μ A 0.017 % 0.025 % 0.012 % 0.02 % 0.13 % 0.1 % 0.30 %	
	15. AC current sourcing instruments Lab-In-House procedure UNI-E003 Rev:0	10 μA – 100 μA 3 Hz – 2 kHz	0.40 %
		100 μA – 1 mA 3 Hz – 10 kHz	0.15 %
		1 mA – 10 mA 3 Hz – 30 kHz >30 kHz – 100 kHz	0.17 % 0.64 %

Schedule



Certificate No. : LA-2023-0845-C

Issue No. : 01

Date : 03 August 2023

Page : 8 of 20

Scope / Instruments / Location / Method	Range to be Calibrated / Measured Quantities / Frequency	CALIBRATION AND MEASUREMENT CAPABILITY (CMC*)
<p>16. Frequency sourcing instruments</p> <p>Lab-In-House procedure UNI-E003 Rev:0</p> <p>17. Capacitance sourcing instruments</p> <p>Lab-In-House procedure UNI-E003 Rev:0</p>	<p>10 mA – 100 mA</p> <p>3 Hz – 2 kHz</p> <p>>2 kHz – 10 kHz</p> <p>>10 kHz – 30 kHz</p>	<p>0.3 %</p> <p>0.2 %</p> <p>0.21 %</p>
	<p>100 mA – 1 A</p> <p>3 Hz – 2 kHz</p> <p>>2 kHz – 10 kHz</p> <p>>10 kHz – 30 kHz</p>	<p>0.57 %</p> <p>0.21 %</p> <p>0.2 %</p>
	<p>1 A – 10 A</p> <p>10 Hz – 2 kHz</p> <p>>2 kHz – 10 kHz</p>	<p>0.2 %</p> <p>0.2 %</p>
	<p>10 A – 30 A</p> <p>10 Hz – 2 kHz</p> <p>>2 kHz – 10 kHz</p>	<p>0.19 %</p> <p>0.19 %</p>
	<p>1 Hz – 10 Hz</p> <p>>10 Hz – 1 kHz</p> <p>>1 kHz – 10 kHz</p> <p>>10 kHz – 100 kHz</p> <p>>100 kHz – 1 MHz</p> <p>>1 MHz – 10 MHz</p> <p>>10 MHz – 100 MHz</p>	<p>0.006 %</p> <p>0.03 %</p> <p>0.0003 %</p> <p>0.0002 %</p> <p>0.001 %</p> <p>0.001 %</p> <p>0.0002 %</p>
	<p>0.1 nF – 1.2 nF</p> <p>>1.2 nF – 12 nF</p> <p>>12 nF – 120 nF</p> <p>>120 nF – 1.2 μF</p> <p>>1.2 μF – 12 μF</p> <p>>12 μF – 120 μF</p> <p>>120 μF – 1.2 mF</p> <p>>1.2 mF – 12 mF</p> <p>>12 mF – 100 mF</p>	<p>2.4 %</p> <p>0.4 %</p> <p>0.2 %</p> <p>0.17 %</p> <p>0.76 %</p> <p>0.3 %</p> <p>0.1 %</p> <p>0.2 %</p> <p>0.1 %</p>

Schedule



Certificate No. : LA-2023-0845-C

Issue No. : 01

Date : 03 August 2023

Page : 9 of 20

Scope / Instruments / Location / Method	Range to be Calibrated / Measured Quantities / Frequency	CALIBRATION AND MEASUREMENT CAPABILITY (CMC*)	
18. DC Power measuring instruments Lab-In-House procedure UNI-E001 Rev:0	0.02 W – 0.5 W	0.74 %	
	>0.5 W – 5 W	0.13%	
	>5 W – 50 W	0.13 %	
	>50 W – 1000 W	0.22 %	
	>1 kW – 5 kW	0.33 %	
	>5 kW – 20 kW	0.34 %	
	>20 kW – 30 kW With Coil 30 kW – 1500 kW	0.33 % 1.0 %	
19. AC Power measuring instruments Lab-In-House procedure UNI-E001 Rev:0	0.02 W – 0.5 W	0.60 %	
	>0.5 W – 5 W	0.14%	
	>5 W – 50 W	0.23 %	
	>50 W – 1000 W	0.7 %	
	>1 kW – 5 kW	2.3 %	
	>5 kW – 20 kW	2.3 %	
	>20 kW – 30 kW With Coil 30 kW – 1500 kW	3 % 4 %	
20. Low Resistance (Earth resistance) – measuring instruments Lab-In-House procedure UNI-E004 Rev:0	100 mΩ – 0.199 Ω	0.013 Ω	
	0.2 Ω – 0.499 Ω	0.013 Ω	
	0.5 Ω – 1.999 Ω	0.051 Ω	
	2 Ω – 4.99 Ω	0.019 Ω	
	5 Ω – 29.9 Ω	0.026 Ω	
	30 Ω – 199.9 Ω	0.12 Ω	
	200 Ω – 499.99 Ω	0.49 Ω	
	500 Ω – 1.999 kΩ	0.0014 kΩ	
	2 kΩ – 4.99 kΩ	0.005 kΩ	
	5 kΩ – 10 kΩ	0.014 kΩ	
	High Resistance (Insulation resistance) – measuring instruments	10 kΩ – 19.999 kΩ	0.025 kΩ
		20 kΩ – 39.99 kΩ	0.047 kΩ
		40 kΩ – 99.99 kΩ	0.12 kΩ
		100 kΩ – 199.99 kΩ	0.25 kΩ
		200 kΩ – 999.9 kΩ	0.47 kΩ
1 MΩ – 1.999 MΩ		0.0081 MΩ	
2 MΩ – 9.999 MΩ		0.01 MΩ	

Schedule



Certificate No. : LA-2023-0845-C

Issue No. : 01

Date : 03 August 2023

Page : 10 of 20

Scope / Instruments / Location / Method	Range to be Calibrated / Measured Quantities / Frequency	CALIBRATION AND MEASUREMENT CAPABILITY (CMC*)
High Resistance (Insulation resistance source with R multiplier) – measuring Instruments	10 MΩ – 19.999 MΩ	0.092 MΩ
	20 MΩ – 199.99 MΩ	0.60 MΩ
	200 MΩ – 999.9 MΩ	1.4 MΩ
	1 GΩ – 1.9 GΩ	0.074 GΩ
	2 GΩ – 10 GΩ	0.077 GΩ
	100 GΩ	0.15 GΩ
	100 GΩ – 999.9 GΩ	2.5 GΩ
21. Ground Bond Resistance measuring instruments (fixed points) Lab-In-House procedure UNI-E004 Rev:0	30 mΩ 55 mΩ 97 mΩ 350 mΩ 490 mΩ 900 mΩ 1.7 Ω 4.7 Ω 9 Ω 17 Ω 47 Ω 90 Ω 170 Ω 470 Ω 900 Ω 1.7 kΩ	1.3 mΩ 2.1 mΩ 2.3 mΩ 4.1 mΩ 4.4 mΩ 9.2 mΩ 0.1 Ω 0.1 Ω 0.1 Ω 0.1 Ω 0.4 Ω 0.6 Ω 1.2 Ω 3.0 Ω 5.9 Ω 0.08 kΩ
22. Leakage Current measuring instruments Lab-In-House procedure UNI-E004 Rev:0	0.1 mA – 30 mA	1.7 %
23. Residual Current measuring instruments Lab-In-House procedure UNI-E004 Rev:0		
Trip current:	3 mA – 3000 mA	1.8 %

Schedule



Certificate No. : LA-2023-0845-C

Issue No. : 01

Date : 03 August 2023

Page : 11 of 20

Scope / Instruments / Location / Method	Range to be Calibrated / Measured Quantities / Frequency	CALIBRATION AND MEASUREMENT CAPABILITY (CMC*)
<p>24. Oscilloscope Calibration</p> <p>Lab-In-House procedure UNI-E005 Rev:0</p> <p>Vertical Deflection @ DC Signal</p>	<p>50 Ω Input</p> <p>± 1 mV – ± 24.999 mV ± 25 mV – ± 109.99 mV ± 110 mV – ± 2.1999 V ± 2.2 V – ± 6.6 V</p> <p>1 MΩ Input</p> <p>± 1 mV – ± 24.999 mV ± 25 mV – ± 109.99 mV ± 110 mV – ± 2.1999 V ± 2.2 – ± 10.999 V ± 11 V – ± 130 V</p>	<p>2 % 0.6 % 6.4 % 0.5 %</p> <p>3.8 % 0.49 % 0.47% 0.46 % 0.46 %</p>
<p>Vertical Deflection @ AC Signal</p>	<p>50 Ω Input</p> <p>1 mVp-p to ± 24.999 mVp-p 25 mV p-p – ±109.99mVp-p 110 mVp-p – ± 2.1999 Vp-p 2.2 Vp-p – ± 6.6 Vp-p</p> <p>1 MΩ Input</p> <p>± 1 mVp-p – ± 24.999 mVp-p ± 25 mV – ± 109.99 mVp-p ± 110 mVp-p – ± 2.1999 Vp-p ± 2.2 Vp-p – ± 10.999 Vp-p ± 11 Vp-p – ± 130 Vp-p</p>	<p>1.9 % 0.6 % 5.3 % 0.5 %</p> <p>3.2 % 0.51 % 0.48% 0.47 % 0.47 %</p>
<p>Horizontal Deflection</p>	<p>2 ns – 5 ns >5 ns – 10 ns >10 ns – 20 ns >20 ns – 50 ns >50 ns – 50 μs >50 μs – 20 ms >20 ms – 50 ms >50 ms – 5 s</p>	<p>0.063 % 0.081 % 0.073 % 0.061 % 0.095 % 0.073 % 0.061 % 0.14%</p>

Schedule



Certificate No. : LA-2023-0845-C

Issue No. : 01

Date : 03 August 2023

Page : 12 of 20

Scope / Instruments / Location / Method	Range to be Calibrated / Measured Quantities / Frequency	CALIBRATION AND MEASUREMENT CAPABILITY (CMC*)
Bandwidth	50 kHz – 100 MHz	0.26 dB
	>100 MHz – 300 MHz	0.30 dB
	>300 MHz – 600 MHz	0.48 dB
	>600 MHz – 1100 MHz	0.57 dB
25. Calibration of Stop Watch Timer	1 s – 60 s	0.25 s
	61 s – 600 s	0.13 s
Lab In-House procedure UNI-E006 Rev:0	601 s – 1200 s	0.13 s
	1201 s – 1800 s	0.13 s
	1801 s – 3600 s	0.13 s
	3601 s – 7200 s	0.13 s
	7201 – 10800 s	0.13 s
B. TEMPERATURE		
1. a) RTD ,PRT Sensor with Indicator & Temperature Measuring instruments (Lab)		0.011 °C
	-80 °C – -40 °C	
	>-40 °C – 0 °C	0.01 °C
	>0 °C – 50 °C	0.01 °C
Lab In-House procedure UNI-T001 Rev:0	>50 °C – 140 °C	0.01 °C
	>140 °C – 210 °C	0.01 °C
	>210 °C – 270 °C	0.011 °C
	>270 °C – 450 °C	0.018 °C
	>450 °C – 500 °C	0.02 °C
	>500 °C – 660 °C	0.39 °C
(Site)		
	-95 °C – 0 °C	0.066 °C
	>0 °C – 60 °C	0.033 °C
	>60 °C – 140 °C	0.047 °C
	>140 °C – 270 °C	0.26 °C
	>270 °C – 450 °C	0.33 °C
	>450 °C – 660 °C	0.39 °C

Schedule



Certificate No. : LA-2023-0845-C

Issue No. : 01

Date : 03 August 2023

Page : 13 of 20

Scope / Instruments / Location / Method	Range to be Calibrated / Measured Quantities / Frequency	CALIBRATION AND MEASUREMENT CAPABILITY (CMC*)
b) Thermocouple Wire with Indicator (Lab)	-80 °C - -40 °C	0.17 °C
	>-40 °C – 0 °C	0.17 °C
	>0 °C – 50 °C	0.17 °C
	>50 °C – 140 °C	0.17 °C
	>140 °C - -210 °C	0.17 °C
	>210 °C – 270 °C	0.17 °C
	>270 °C – 450 °C	0.17 °C
	>450 °C – 500 °C	0.17 °C
	>500 °C – 660 °C	0.53 °C
	>660 °C – 800 °C	0.9 °C
	>800 °C – 1000 °C	0.9 °C
(Site)	-95 °C – 0 °C	0.29 °C
	>0 °C – 60 °C	0.24 °C
	>60 °C – 140 °C	0.24 °C
	>140 °C – 270 °C	0.31 °C
	>270 °C – 450 °C	0.33 °C
	>450 °C – 660 °C	0.4 °C
(C) Temperature Indicator with other Sensors (Lab)	-80 °C – 0 °C	0.14 °C
	>0 °C – 50 °C	0.12 °C
	>50 °C – 140 °C	0.12 °C
	>140 °C – 210 °C	0.13 °C
	>210 °C – 270 °C	0.13 °C
	>270 °C – 450 °C	0.13 °C
	>450 °C – 500 °C	0.13 °C
>500 °C – 660 °C	0.53 °C	
(Site)	-95 °C – 0 °C	0.17 °C
	>0 °C – 60 °C	0.13 °C
	>60 °C – 140 °C	0.13 °C
	>140 °C – 270 °C	0.28 °C
	>270 °C – 450 °C	0.35 °C
	>450 °C – 660 °C	0.4 °C

Schedule



Certificate No. : LA-2023-0845-C

Issue No. : 01

Date : 03 August 2023

Page : 14 of 20

Scope / Instruments / Location / Method	Range to be Calibrated / Measured Quantities / Frequency	CALIBRATION AND MEASUREMENT CAPABILITY (CMC*)
2. Fixed Point RTD Sensor Lab In-House procedure UNI-T002 Rev:0	Approx. -196.0 °C 0.01 (TPW) 29.7647 (Gallium) 660.3226 (Aluminium)	0.012 °C 3 mK 3 mK 9 mK
Thermistor	0.01 (TPW) 29.7647 (Gallium)	3 mK 3 mK
3. RTD Sensor without Indicator, PRT Sensor without Indicator & Temperature measuring instruments (Lab) Lab In-House procedure UNI-T003 Rev:0	-80 °C - -40 °C >-40 °C - 0 °C >0 °C - 50 °C >50 °C - 140 °C >140 °C - 210 °C >210 °C - 270 °C >270 °C - 500 °C >500 °C - 660 °C	0.01 °C 0.01 °C 0.01 °C 0.012 °C 0.012 °C 0.02 °C 0.03 °C 0.39 °C
(Site)	-95 °C - 60 °C >60 °C - 140 °C >140 °C - 270 °C >270 °C - 450 °C >450 °C - 660 °C	0.6 °C 0.35 °C 0.43 °C 0.47 °C 0.52 °C
4. Thermocouple Wire without Indicator (Lab) Lab In-House procedure UNI-T004 Rev:0	-80 °C - -40 °C >-40 °C - 0 °C >0 °C - 50 °C >50 °C - 140 °C >140 °C - 210 °C >210 °C - 270 °C >270 °C - 450 °C >450 °C - 500 °C >500 °C - 660 °C >660 °C - 800 °C >800 °C - 1000 °C	0.8 °C 0.8 °C 0.36 °C 0.36 °C 0.36 °C 0.36 °C 0.36 °C 0.36 °C 0.36 °C 0.5 °C 1.0 °C 1.1 °C

Schedule



Certificate No. : LA-2023-0845-C

Issue No. : 01

Date : 03 August 2023

Page : 15 of 20

Scope / Instruments / Location / Method	Range to be Calibrated / Measured Quantities / Frequency	CALIBRATION AND MEASUREMENT CAPABILITY (CMC*)
(Site)	-95 °C – 60 °C >60 °C – 140 °C >140 °C – 280 °C >280 °C – 450 °C >450 °C – 660 °C	0.72 °C 0.71 °C 0.77 °C 0.78 °C 0.8 °C
4 (a) Connecting Thermocouple wire		
Type S	0 °C – 30 °C	0.0003 mV
Type R		0.0003 mV
Type J		0.001 mV
Type K		0.0008 mV
Type N		0.0005 mV
Type T		0.0008 mV
Type E		0.0012 mV
(b) C/J Thermocouple Wire		
Type S	0 °C – 30 °C	0.0002 mV
Type R		0.0002 mV
Type J		0.0007 mV
Type K		0.0006 mV
Type N		0.0006 mV
Type T		0.0006 mV
Type E		0.0008 mV

Schedule



Certificate No. : LA-2023-0845-C

Issue No. : 01

Date : 03 August 2023

Page : 16 of 20

Scope / Instruments / Location / Method	Range to be Calibrated / Measured Quantities / Frequency	CALIBRATION AND MEASUREMENT CAPABILITY (CMC*)														
5. Thermocouple Indicator Lab In-House procedure UNI-T005 Rev:0 Type K Type: N Type: J Type: T Type: R	-200 °C - -100 °C >-100 °C - -25 °C >-25 °C – 120 °C >120 °C – 600 °C >600 °C – 1000 °C >1000 °C – 1372 °C	<table border="1"> <thead> <tr> <th>Lab</th> <th>Site</th> </tr> </thead> <tbody> <tr><td>0.51 °C</td><td>0.91 °C</td></tr> <tr><td>0.43 °C</td><td>0.55 °C</td></tr> <tr><td>0.46 °C</td><td>0.55 °C</td></tr> <tr><td>0.46 °C</td><td>0.55 °C</td></tr> <tr><td>0.56 °C</td><td>0.72 °C</td></tr> <tr><td>0.56 °C</td><td>0.91 °C</td></tr> </tbody> </table>	Lab	Site	0.51 °C	0.91 °C	0.43 °C	0.55 °C	0.46 °C	0.55 °C	0.46 °C	0.55 °C	0.56 °C	0.72 °C	0.56 °C	0.91 °C
	Lab	Site														
	0.51 °C	0.91 °C														
	0.43 °C	0.55 °C														
	0.46 °C	0.55 °C														
	0.46 °C	0.55 °C														
	0.56 °C	0.72 °C														
	0.56 °C	0.91 °C														
	-200 °C - -100 °C >-100 °C - -25 °C >-25 °C – 120 °C >120 °C – 410 °C >410 °C – 1000 °C >1000 °C – 1300 °C	<table border="1"> <thead> <tr> <th>Lab</th> <th>Site</th> </tr> </thead> <tbody> <tr><td>0.68 °C</td><td>0.68 °C</td></tr> <tr><td>0.59 °C</td><td>0.82 °C</td></tr> <tr><td>0.58 °C</td><td>0.82 °C</td></tr> <tr><td>0.61 °C</td><td>0.82 °C</td></tr> <tr><td>0.61 °C</td><td>0.82 °C</td></tr> <tr><td>0.61 °C</td><td>0.91 °C</td></tr> </tbody> </table>	Lab	Site	0.68 °C	0.68 °C	0.59 °C	0.82 °C	0.58 °C	0.82 °C	0.61 °C	0.82 °C	0.61 °C	0.82 °C	0.61 °C	0.91 °C
	Lab	Site														
	0.68 °C	0.68 °C														
	0.59 °C	0.82 °C														
	0.58 °C	0.82 °C														
	0.61 °C	0.82 °C														
	0.61 °C	0.82 °C														
	0.61 °C	0.91 °C														
	210 - -100 °C >-100 °C - -30 °C >-30 °C – 150 °C >150 °C – 760 °C >760 °C – 1000 °C >1000 °C – 1200 °C	<table border="1"> <thead> <tr> <th>Lab</th> <th>Site</th> </tr> </thead> <tbody> <tr><td>0.57 °C</td><td>0.91 °C</td></tr> <tr><td>0.51 °C</td><td>0.64 °C</td></tr> <tr><td>0.52 °C</td><td>0.64 °C</td></tr> <tr><td>0.54 °C</td><td>0.64 °C</td></tr> <tr><td>0.54 °C</td><td>0.81 °C</td></tr> <tr><td>0.54 °C</td><td>0.81 °C</td></tr> </tbody> </table>	Lab	Site	0.57 °C	0.91 °C	0.51 °C	0.64 °C	0.52 °C	0.64 °C	0.54 °C	0.64 °C	0.54 °C	0.81 °C	0.54 °C	0.81 °C
	Lab	Site														
	0.57 °C	0.91 °C														
	0.51 °C	0.64 °C														
	0.52 °C	0.64 °C														
	0.54 °C	0.64 °C														
	0.54 °C	0.81 °C														
	0.54 °C	0.81 °C														
-250 °C – -150 °C >-150 °C – 0 °C >0 °C – 120 °C >120 °C – 400 °C	<table border="1"> <thead> <tr> <th>Lab</th> <th>Site</th> </tr> </thead> <tbody> <tr><td>0.86 °C</td><td>2.2 °C</td></tr> <tr><td>0.47 °C</td><td>0.55 °C</td></tr> <tr><td>0.46 °C</td><td>0.55 °C</td></tr> <tr><td>0.46 °C</td><td>0.55 °C</td></tr> </tbody> </table>	Lab	Site	0.86 °C	2.2 °C	0.47 °C	0.55 °C	0.46 °C	0.55 °C	0.46 °C	0.55 °C					
Lab	Site															
0.86 °C	2.2 °C															
0.47 °C	0.55 °C															
0.46 °C	0.55 °C															
0.46 °C	0.55 °C															
0 °C – 250 °C >250 °C – 400 °C >400 °C – 1000 °C >1000 °C – 1770 °C	<table border="1"> <thead> <tr> <th>Lab</th> <th>Site</th> </tr> </thead> <tbody> <tr><td>0.91 °C</td><td>2.0 °C</td></tr> <tr><td>0.72 °C</td><td>1.5 °C</td></tr> <tr><td>0.76 °C</td><td>1.5 °C</td></tr> <tr><td>0.76 °C</td><td>1.5 °C</td></tr> </tbody> </table>	Lab	Site	0.91 °C	2.0 °C	0.72 °C	1.5 °C	0.76 °C	1.5 °C	0.76 °C	1.5 °C					
Lab	Site															
0.91 °C	2.0 °C															
0.72 °C	1.5 °C															
0.76 °C	1.5 °C															
0.76 °C	1.5 °C															

Schedule



Certificate No. : LA-2023-0845-C

Issue No. : 01

Date : 03 August 2023

Page : 17 of 20

Scope / Instruments / Location / Method	Range to be Calibrated / Measured Quantities / Frequency	CALIBRATION AND MEASUREMENT CAPABILITY (CMC*)	
		(Lab)	(Site)
Type: S	0 °C – 250 °C >250 °C – 1000 °C >1000 °C – 1400 °C >1400 °C – 1760 °C	0.82 °C 0.72 °C 0.79 °C 0.79 °C	2.0 °C 1.3 °C 1.3 °C 1.6 °C
Type: E	-250 °C - -150 °C >-150 °C - -25 °C >-25 °C – 350 °C >350 °C – 650 °C >650 °C – 1000 °C	0.66 °C 0.48 °C 0.49 °C 0.52 °C 0.52 °C	1.7 °C 0.6 °C 0.6 °C 0.68 °C 0.68 °C
6. Thermocouple Simulator Lab In-House procedure UNI-T005 Rev:0		(Lab)	(Site)
Type: K	-200 °C - -100 °C >-100 °C - -25 °C >25 °C – 120 °C >120 °C – 600 °C >600 °C – 1000 °C >1000 °C – 1372 °C	0.51 °C 0.42 °C 0.46 °C 0.46 °C 0.56 °C 0.56 °C	0.62 °C 0.55 °C 0.55 °C 0.55 °C 0.55 °C 0.55 °C
Type: N	-200 °C - -100 °C >-100 °C - -25 °C >-25 °C – 120 °C >120 °C – 410 °C >410 °C – 1000 °C >1000 °C – 1300 °C	0.68 °C 0.58 °C 0.58 °C 0.61 °C 0.61 °C 0.61 °C	0.91 °C 0.82 °C 0.82 °C 0.82 °C 0.68 °C 0.68 °C
Type: J	210 °C - -100 °C >-100 °C - -30 °C >-30 °C – 150 °C >150 °C – 760 °C >760 °C – 1000 °C >1000 °C – 1200 °C	0.57 °C 0.51 °C 0.52 °C 0.52 °C 0.54 °C 0.54 °C	0.63 °C 0.57 °C 0.57 °C 0.57 °C 0.63 °C 0.63 °C

Schedule



Certificate No. : LA-2023-0845-C

Issue No. : 01

Date : 03 August 2023

Page : 18 of 20

Scope / Instruments / Location / Method	Range to be Calibrated / Measured Quantities / Frequency	CALIBRATION AND MEASUREMENT CAPABILITY (CMC*)	
		(Lab)	(Site)
Type: T	-250 °C - -150 °C >-150 °C – 0 °C >0 °C – 120 °C >120 °C – 400 °C	0.86 °C 0.50 °C 0.46 °C 0.46 °C	1.3 °C 0.61 °C 0.61 °C 0.61 °C
Type: R	0 °C – 250 °C >250 °C – 400 °C >400 °C – 1000 °C >1000 °C – 1770 °C	0.91 °C 0.72 °C 0.67 °C 0.72 °C	1.6 °C 1.4 °C 1.4 °C 1.4 °C
Type: S	0 °C – 250 °C >250 °C – 1000 °C >1000 °C – 1400 °C >1400 °C – 1760 °C	0.79 °C 0.72 °C 0.76 °C 0.76 °C	1.6 °C 1.3 °C 1.4 °C 1.3 °C
Type: E	-250 °C - -150 °C >-150 °C - -25 °C >-25 °C – 350 °C >350 °C – 650 °C >650 °C – 1000 °C	0.65 °C 0.45 °C 0.47 °C 0.56 °C 0.51 °C	0.91 °C 0.64 °C 0.64 °C 0.57 °C 0.57 °C
7. RTD and PRT Indicator			
Lab In-House procedure UNI-T005 Rev:0	-200 °C - -80 °C >-80 – 0 °C >0 °C – 100 °C	0.06 °C 0.06 °C 0.07 °C	0.1 °C 0.1 °C 0.1 °C
(Lab / Site)	>100 °C – 300 °C >300 °C – 400 °C	0.1 °C 0.11 °C	0.15 °C 0.17 °C
TYPE: PT 385 (100Ω)	>400 °C – 630 °C >630 °C – 800 °C	0.14 °C 0.23 °C	0.22 °C 0.25 °C
8. RTD Simulator			
Lab In-House procedure UNI-T006 Rev:0	-200 °C – 80 °C >80 °C – 100 °C >100 °C – 200 °C	0.02 °C 0.01 °C 0.01 °C	0.15 °C 0.15 °C 0.16 °C
(Lab / Site)	>200 °C – 400 °C >400 °C – 800 °C	0.01 °C 0.01 °C	0.27 °C 0.27 °C
TYPE: PT 385 (100Ω)	-		

Schedule



Certificate No. : LA-2023-0845-C

Issue No. : 01

Date : 03 August 2023

Page : 19 of 20

Scope / Instruments / Location / Method	Range to be Calibrated / Measured Quantities / Frequency	CALIBRATION AND MEASUREMENT CAPABILITY (CMC*)
<p>9. Dry Block Calibrator</p> <p>Lab In-House procedure UNI-T007 Rev:0</p> <p>(Lab / Site)</p>	<p>-95 °C – 140 °C >140 °C – 450 °C >450 °C – 600 °C</p>	<p>0.035 °C 0.05 °C 0.052 °C</p>
<p>10. Non-Contact Temperature Infrared Thermometer / Thermal imager</p> <p>Lab In-House procedure UNI-T008 Rev:0</p> <p>Emissivity: 0.90 – 1.00</p> <p>Spectral Band: 8 to 14 µm</p>	<p>35 °C – 100 °C >100 °C – 200 °C >200 °C – 350 °C >350 °C – 500 °C</p>	<p>1.1 °C 2.1 °C 2.6 °C 3.7 °C</p>
<p>11. Humidity / Temperature measuring instruments</p> <p>Lab In-House procedure UNI-T009 Rev:0</p> <p>(Lab / Site) (Used Reference Probe)</p> <p>Temperature:</p> <p>Humidity:</p> <p>Lab : (Used Chilled Mirror) Temperature</p> <p>Humidity</p> <p>Fixed Temperature @23°C</p>	<p>5 °C – 50 °C</p> <p>(5 – 90) % relative humidity</p> <p>15 °C – 50 °C</p> <p>(40 – 90) % relative humidity</p> <p>(05 – 50) % relative humidity (50 – 90) % relative humidity</p>	<p>0.3 °C</p> <p>2.4 % relative humidity</p> <p>0.2 °C</p> <p>1.2 % relative humidity</p> <p>0.7 % relative humidity 1.1 % relative humidity</p>

Schedule



Certificate No. : LA-2023-0845-C

Issue No. : 01

Date : 03 August 2023

Page : 20 of 20

Scope / Instruments / Location / Method	Range to be Calibrated / Measured Quantities / Frequency	CALIBRATION AND MEASUREMENT CAPABILITY (CMC*)
12. Temperature Enclosure		
Lab In-House procedure UNI-T010 Rev:0	-80 °C – -40 °C	0.4 °C
	-40 °C – 50 °C	0.4 °C
	50 °C – 160 °C	0.4 °C
(Lab / Site)	160 °C – 200 °C	1.1 °C
	200 °C – 600 °C	2.6 °C
Freezers Oven & Furnaces	600 °C – 800 °C	2.6 °C
	800 °C – 1000 °C	3.7 °C

Approved Signatories

- Mr Gavin Tan - For Electrical, A1 to A19
- Mr Prabhakar Sundararaj - For Temperature, B5 to B8 and B10
For Electrical, all items in Cat A
- Mr S.Karupiah - For Temperature, all items in Cat B
For Electrical, all items in Cat A

Note :

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025. A laboratory's fulfilment of the requirements of ISO/IEC 17025 means the laboratory meets both the technical competence requirements and **management system requirements** that are necessary for it to consistently deliver technically valid calibration results. The **management system requirements** in ISO/IEC 17025 are written in language relevant to laboratory operations and operate generally in accordance with the principles of ISO 9001.